

University of North Bengal

Tender Notice

Sealed quotations are invited from interested vendors/suppliers for supplying Microscope to the Department of Microbiology, UNB. For specification, terms and conditions and other details, please visit www.nbu.ac.in

Sd/-

Registrar

Advt. No. 223/R-2019

Date: 09.01.2019

Enclosure-I

University of North Bengal
Raja Rammohunpur, Dist-Darjeeling (W.B.)

TENDER PAPER FOR EQUIPMENT

Sl. No.	Item	Specification	Quantity	Rate	Amount (in Rs) inc. GST
1.	Microscope	Annexure-B	01		

It is certified that on behalf of the organization we are quoting the above prices as per the terms and conditions laid by the University of North Bengal in Annexure-A. We also certify that we have branch/service centre at Kolkata/Siliguri and after sales service will be provided within 48hrs of call booking.

Name of the company:

Signature:

Address:

ANNEXURE-A

TERMS AND CONDITION FOR EQUIPMENT

- (1) The quoted rate shall be inclusive of all taxes.**
- (2) The equipment shall carry three years onsite warranty from the date of delivery.**
- (3) The equipment should be as per specification.**
- (4) Taxes will be deducted at source as per prevailing rules of Central and State Government.**
- (5) Bill will be released after delivery and successful installation of the equipment(s) subject to satisfactory performance report by the Head of the Department.**
- (6) The company must have service centre at Kolkata/Siliguri.**
- (7) Copy of PT, Trade License, PAN card and GST duly signed and stamped by the vendor shall be enclosed along with the tender.**
- (8) The service engineer shall attend the call within 48 hrs for trouble shooting to be done on no wait basis.**
- (9) The successful tenderer shall supply the equipment at the Department of Microbiology, University of North Bengal, within 6 to 8 weeks from the date of issue of the supply order.**
- (10) Tender papers maybe downloaded from NBU website (www.nbu.ac.in).**
- (11) The University authority reserves the right to accept or rejects any/all quotations without assigning any reason and shall not be bound to accept the lowest quotation.**
- (12) The quotation should be valid for at least three months.**
- (13) The brochure /catalogue of the equipment shall be enclosed with the tender.**
- (14) Selection of the agency will be made on the basis of both technical and financial bids. Technical bids and financial bids shall be sealed by the bidder in separate covers duly superscribed and both the sealed covers shall be put in a bigger cover which should also be sealed and duly superscribed. Financial bids of only technically qualified bidders will be considered.**
- (15) The last date of submission of tender form is 7 working days from the date of publication of advertisement and to be opened on the same date in the Dept of Microbiology, UNB.**
- (16) For any clarification regarding tender please contact with HoD (Ph.No. 0353-2776319)**
- (17) The duly filled up tender paper is to be sent to the under noted address.**

**Head
Department of Microbiology
University of North Bengal
Raja Rammohunpur, P.O. NBU
Dist: Darjeeling. PIN-734013**

Sd/-

Registrar

ANNEXURE-B

Specification for Microscope:

1. Inclination angle adjustable to 30 degree/45 degree
2. Standing position with 360 degree rotation
3. Optics should be hard coated & Anti-fungus treated
4. Observation head & all optics should be Multi-layer coated
5. Should be supplied with ball bearing Reverse mounted nosepiece
6. Co-axial Coarse & Fine focus on friction bearing should be present
7. Co-axial Mechanical stage with low position controls for easy access should be present
8. Abbe Condenser N.A 0.9/DRY/1.25 OIL with iris diaphragm should be present
9. Should be supplied with Halogen Koehler's illumination 30W or LED Illumination with battery back up
10. Complete dust cover should be present
11. Should be supplied with Infinity corrected optics
12. Wide field eyepieces (paired) WF-10X should be present
13. Objectives (5 No's) 10X, 20X, 40X , 60X & 100X- oil immersion (Anti fungus) should be present
14. Warranty should be 3 years or more
15. Should be supplied with Dust cover, Spare lamp & working manual

Specification for Camera:

1. To be very convenient to use, just to insert into the eyepiece tube or top tube of Trinocular head
2. The Visual Field observed from the Camera should be the Viewing Field from the eyepiece
3. By USB 2.0 getting Real-time, Non-compressing Video data and Image capturing directly should be present
4. High sensitivity features suitable for wide range of accessories such as Dark field. Phase contrast & Epi-Fluorescence should be present
5. Should be supplied with NAVIGATION SOFTWARE: For sorting, retrieving, E-Mailing images including various functions like zooming and image contrast etc