

#### University of North Bengal Office of The Registrar Accredited by NAAC with Grade B<sup>++</sup> Tender Notice

Tenders are invited from interested vendors/suppliers for supplying Shaking Incubator to the Department of Microbiology, NBU. For specification, terms and conditions and other details, please visit <u>www.nbu.ac.in</u>

Sd/-

Registrar

Advt. No. 149/R-2023

Date: 28.03.2023



## University of North Bengal Office of The Registrar Accredited by NAAC with Grade B<sup>++</sup> TENDER PAPER FOR EQUIPMENT

| Sl.<br>No. | Item                 | Specification | Quantity | Rate | Discount | Amount (in<br>Rs) inc. GST |
|------------|----------------------|---------------|----------|------|----------|----------------------------|
| 1.         | Shaking<br>Incubator | Annexure-B    | 01       |      |          |                            |

It is certified that on behalf of the organization we are quoting the above prices as per the terms and conditions laid by the University of North Bengal in Annexure-A. We also certify that we have branch/service centre at Kolkata and after sales service will be provided within 48hrs of call booking.

Name of the company:

Signature:

Address:

## **ANNEXURE-A**

#### TERMS AND CONDITION FOR EQUIPMENT

- (1) The quoted rate shall be inclusive of all taxes.
- (2) The equipment shall carry three years of onsite warranty from the date of delivery.
- (3) The equipment should be as per specification.
- (4) Taxes will be deducted at source as per prevailing rules of Central and State Government.
- (5) Bill will be released after delivery and successful installation of the equipment(s) subject to satisfactory performance report by the Head of the Department.
- (6) The company must have service centre at Kolkata.
- (7) Copy of PT, Trade License and PAN card and IT and GST duly signed and stamped by the vendor shall be enclosed along with the tender. The vendor shall have a valid GST registration number and the number of same be mentioned.
- (8) The service engineer shall attend the call within 48 hrs for troubleshooting to be done on no wait basis.
- (9) The successful tenderer shall supply the equipment at the Department of Biotechnology, University of North Bengal, within 4 weeks from the date of issue of the supply order.
- (10) Tender papers maybe downloaded from NBU website (www.nbu.ac.in).
- (11) The University authority reserves the right to accept or rejects any/all quotations without assigning any reason and shall not be bound to accept the lowest quotation.
- (12) The quotation should be valid for atleast three months.
- (13) The brochure /catalogue of the equipment shall be enclosed with the tender.
- (14) Selection of the agency will be made on the basis of both technical and financial bids. Technical bids and financial bids shall be sealed by the bidder in separate covers duly superscribed and both the sealed covers shall be put in bigger cover which should also be sealed and duly superscribed. Financial bids of only technically qualified bidders will be considered.
- (15) The last date of submission of tender form is **31.03.2023** and to be opened on same day in the Dept of Microbiology, NBU.
- (16) For any clarification regarding tender please contact with HoD.
- (17) The duly filled up tender paper is to be sent to the under noted address.

Khusboo Lepcha Head Department of Microbiology University of North Bengal Raja Rammohunpur, P.O. NBU Dist: Darjeeling. PIN-734013

# **Annexure-B**

**Incubator with Orbital Shaker Mechanism Exterior Material** : MS Powder coated **Interior Material** : SS 304 Temperature Range : Ambient+5 to 60 °C Temperature Controller : Digital display with microprocessor based PID Controller with PT 100 sensor Accuracy : ±0.5 °C Samples viewing: Transparent glass window cut out on front door of 1 Sq Ft **In-built lighting:** Built-in LED lighting provided with external switch on front panel **RPM Shaking speed** : 20 to 200 RPM without load **Uniformity Control** : Recirculation fan inside for uniformity of temperature Mains Power : Works on 230 V AC single phase 50Hz Tray size-325x325mm **Chamber Size** L x D x H: 400 x 425 x 450mm Stroke Length: 25mm dia

Stroke Speed per min: 20 to 200